

XYZ Pharmaceutical (Pvt) Ltd.
QUALITY CONTROL DEPARTMENT

TITLE:

CALIBRATION REPORT OF DISSOLUTION TESTER

Model No. -----

Manufacturer: -----

Instrument ID: -----

Environmental Details:

Temperature: -----

Relative Humidity: -----

Calibration Standards:

Tachometer

Pre-calibrated Digital stopwatch

Pre-calibrated Digital thermometer

Calibration Date:	Last Calibration done:	Next Due Calibration:
--------------------------	-------------------------------	------------------------------

Temperature

S. No.	Time	Set Temp	Mean Observed Temp	Deviation	Limits	Status	Performed by
1	After 1 hour	37°C			36.5 to 37.5°C		
2	After next 1 hour	37°C			36.5 to 37.5°C		

RPM

S. No.	Set RPM	Mean Observed RPM	Deviation	Limits	Status	Performed by
1	50			± 4%		
2	75			± 4%		
3	100			± 4%		

Time

S. No.	Set Time	Mean Observed Time	Deviation	Limits	Status	Performed by
1	30 mint					
2	60 mint					
3	120 mint					