

XYZ Pharmaceutical (Pvt) Ltd.
QUALITY CONTROL DEPARTMENT

TITLE:

CALIBRATION REPORT OF REFRACTOMETER

Model No. -----

Manufacturer: -----

Instrument ID: -----

Environmental Details:

Temperature: -----

Relative Humidity: -----

Calibration Standards:

Carbon tetrachloride: 1.4603

Toluene: 1.4969

Water: 1.3325

Calibration Date:

Last Calibration done:

Next Due Calibration:

S. No.	Reference Liquids	Refractive Index at 20 °C	Observed Reading	Deviation	Limits	Status	Performed by
1	Water	1.3325			± 0.0001		
2	Carbon tetra chloride	1.4603			± 1		
3	Toluene	1.4969			± 1		