

XYZ Pharmaceutical (Pvt) Ltd.
QUALITY CONTROL DEPARTMENT

TITLE:

CALIBRATION REPORT OF UV-VIS SPECTROMETER

Model No. -----

Manufacturer: -----

Instrument ID: -----

Environmental Details:

Temperature: -----

Relative Humidity: -----

Calibration Standards:

Potassium Dichromate Solution

Calibration Date:	Last Calibration done:	Next Due Calibration:
--------------------------	-------------------------------	------------------------------

Wavelength (nm)	A (1% 1cm)	Limit
235		122.9 – 126.2
257		142.8 – 145.7
313		47.0 – 50.3
350		105.3 – 108.2